

EPA Protozoan Proficiency Testing Laboratory Results Summary Round 0606

Laboratory Name: BioVir Laboratories, Inc
Contact Name: Dr. Richard Danielson
Address:
 685 Stone Road, Unit #6
 Benicia, CA 94510-1126

Laboratory ID: CA01401
Filter: Envirochek HV
Volume Spiked: 100 L
Elution: Wrist Shaker
Concentration: Centrifugation

Sample Date	Sample ID	Turbidity (NTU) Reported	<i>Cryptosporidium</i>			<i>Giardia</i>		
			Oocysts Reported	Oocysts Spiked	% Recovered	Cysts Reported	Cysts Spiked	% Recovered
6/19/2006	03440	1.65	123	189.90	65	112	201.71	56
6/19/2006	03441	0.5	132	189.90	70	113	201.71	56
6/19/2006	03442	1.1	105	169.71	62	0	0.00	na

Spike Status	Parameter	<i>Cryptosporidium</i>	<i>Giardia</i>
Combo	Mean % Recovery	65	56
	RSD of % Recovery	6	1
	No. of False Positives	0	0