



A Waters Company

Richard Danielson
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510

WS-198



Final Report

WatR™ Supply Proficiency Testing

WatR™ Supply Study

Open Date: 01/07/13

Close Date: 02/21/13

Report Issued Date: 03/08/13



A Waters Company

March 8, 2013

Richard Danielson
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510

Enclosed is your final report for ERA's WS-198 WatR™Supply Proficiency Testing (PT) study. Your final report includes an evaluation of all results submitted by your laboratory to ERA.

Data Evaluation Protocols: All analytes in ERA's WS-198 WatR™Supply Proficiency Testing (PT) study have been evaluated using the following tiered approach. If the analyte is listed in the current TNI Fields of Proficiency Testing (FoPT) tables, the evaluation was completed by comparing the reported result to the acceptance limits generated using the criteria contained in the TNI FoPT tables. If the analyte is not included in the TNI FoPT tables, the reported result has been evaluated using the procedures outlined in ERA's Standard Operating Procedure for the Generation of Performance Acceptance Limits (SOP 0260).

Corrective Action Help: As part of your accreditation(s), you may be required to identify the root cause of any "Not Acceptable" results, implement the necessary corrective actions, and then satisfy your PT requirements by participating in a Supplemental (QuiK™Response) or future ERA PT study. ERA's technical staff is available to help your laboratory resolve any technical issues that may be impairing your PT performance and possibly affecting your routine data quality. Our laboratory and technical staff have many years of collective experience in performing the full range of environmental analyses. As part of our technical support, ERA offers QC samples that can be useful in helping you work through your technical issues.

Please note the following changes to our final reports:

- At the request of the TNI Accreditation Council, we have included a Laboratory Exception Report that includes a list of all analytes reported with less than qualifiers when the assigned value was greater than "0." In addition, because we have received many requests from laboratories, this report also includes a list of all analytes with "Not Acceptable" evaluations.
- Some states have elected not to convert to the 2009 TNI Standards at this time. If you have released your results to a state that has retained the 2003 NELAC Evaluation Criteria, your final report will include a section that evaluates the results according to the 2003 Standard in addition to the 2009 TNI Standards.

Thank you for your participation in ERA's WS-198 WatR™Supply Proficiency Testing study. If you have any questions, please contact our Proficiency Testing Department at 1-800-372-0122.

Sincerely,

A handwritten signature in black ink, appearing to read "Kristina Sanchez", written over a faint circular stamp or watermark.

Kristina Sanchez
Quality Officer

attachments



A Waters Company

Report Recipient	Contact/Phone Number	Reporting Type	Evaluation Type
California	Fred Choske / 510-620-3175	All Analytes	2009 TNI

WS-198 Definitions & Study Discussion

Study Dates: 01/07/13 - 02/21/13

Report Issued: 03/08/13

WS Study Definitions

The Reported Value is the value that the laboratory reported to ERA.

The ERA Assigned Values are compliant with the current TNI Fields of Proficiency Testing (FoPT) tables. A parameter not added to the standard is given an Assigned Value of "< PTRL" per the guidelines contained in the 2009 TNI Standards. The assigned values are directly traceable to the commercially prepared starting materials used to manufacture the PT standards.

The Acceptance Limits are established per the criteria contained in the most current TNI FoPT tables, or ERA's SOP for the Generation of Performance Acceptance Limits™ as applicable.

The Performance Evaluation:

- Acceptable = Reported Value falls within the Acceptance Limits.
- Not Acceptable = Reported Value falls outside the Acceptance Limits.
- No Evaluation = Reported Value cannot be evaluated.
- Not Reported = No Value reported.

The Method Description is the method the laboratory reported to ERA.

WS Study Discussion

ERA's WS-198 WatR™ Supply Proficiency Testing study has been reviewed by ERA senior management and certified compliant with the requirements of the 2009 TNI PT Standards and the criteria contained in the most current TNI Fields of Proficiency Testing (FoPT) tables.

ERA's WS-198 WatR™ Supply study standards were examined for any anomalies. A full review of all homogeneity, stability and accuracy verification data was completed. All analytical verification data for all analytes in the standards met the acceptance criteria contained in the 2009 TNI PT Standard and the criteria contained in the most current TNI FoPT tables.

The data submitted by participating laboratories was also examined for study anomalies. There were no anomalies observed during the statistical review of the data.

ERA's WS-198 WatR™ Supply study reports shall not be reproduced except in their entirety and not without the permission of the participating laboratories. The report must not be used by the participating laboratories to claim product endorsement by any agency of the U. S. government.

The data contained herein are confidential and intended for your use only.

If you have any questions or concerns regarding your assessment in ERA's WatR™ Supply Proficiency Testing program, please contact our Proficiency Testing Department at 1-800-372-0122.





A Waters Company

WS-198 Laboratory Exception Report

Richard Danielson
Lab Director
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510
(707) 747-5906

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01401
B354201
03/08/13
01/07/13 - 02/21/13

2009 TNI Evaluation Checks

There are no values reported with < where the assigned value was greater than 0.

2009 TNI Not Acceptable Evaluations

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description
WS MicrobE™ (Coliforms) (cat# 080A)							
2500	Total Coliforms - Sample 8	CFU/100mL	Absence	Presence	Presence	Not Acceptable	SM9223 COLILERT online
WS MicrobE™ (Coliforms) (cat# 080D)							
2530	Fecal Coliforms - Sample 2	CFU/100mL	Presence	Absence	Absence	Not Acceptable	SM9222D m FC online



All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01



Final Report Results For Laboratory Biovir Laboratories



2009 TNI Evaluation Report

Study: **WS-198**

ERA Customer Number: **B354201**

Laboratory Name: **Biovir Laboratories**

Microbiology Results





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080A)

2500	Total Coliforms - Sample 1	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 2	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 3	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 4	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 5	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 6	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 7	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 8	CFU/100mL	Absence	Presence	Presence	Not Acceptable	SM9223 COLILERT online	2/15/2013				RN
2500	Total Coliforms - Sample 9	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2500	Total Coliforms - Sample 10	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN

Total Coliforms Evaluation : Not Acceptable

Fecal Coliforms Evaluation : Not Reported

E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 3, 5 and 10

Total Coliform Organism - Enterobacter cloacae, Samples 1, 6 and 8

Negative (1) Coliform Organism - Proteus mirabilis, Sample 9

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 2

Blank - No Organism, Samples 4 and 7



All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080A) (Continued)

2530	Fecal Coliforms - Sample 1	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 2	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 3	CFU/100mL		Presence	Presence	Not Reported						
2530	Fecal Coliforms - Sample 4	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 5	CFU/100mL		Presence	Presence	Not Reported						
2530	Fecal Coliforms - Sample 6	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 7	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 8	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 9	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 10	CFU/100mL		Presence	Presence	Not Reported						

Total Coliforms Evaluation : Not Acceptable

Fecal Coliforms Evaluation : Not Reported

E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 3, 5 and 10

Total Coliform Organism - Enterobacter cloacae, Samples 1, 6 and 8

Negative (1) Coliform Organism - Proteus mirabilis, Sample 9

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 2

Blank - No Organism, Samples 4 and 7





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080A) (Continued)

2525	E.coli - Sample 1	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 2	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 3	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 4	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 5	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 6	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 7	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 8	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 9	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN
2525	E.coli - Sample 10	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9223 COLILERT online	2/15/2013	0			RN

Total Coliforms Evaluation : Not Acceptable

Fecal Coliforms Evaluation : Not Reported

E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 3, 5 and 10

Total Coliform Organism - Enterobacter cloacae, Samples 1, 6 and 8

Negative (1) Coliform Organism - Proteus mirabilis, Sample 9

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 2

Blank - No Organism, Samples 4 and 7

All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080B)

2500	Total Coliforms - Sample 1	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 2	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 3	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 4	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 5	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 6	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 7	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 8	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 9	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN
2500	Total Coliforms - Sample 10	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221B 18/19thED	2/15/2013	0			RN

Total Coliforms Evaluation : Acceptable

Fecal Coliforms Evaluation : Acceptable

E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 1, 4 and 6

Total Coliform Organism - Enterobacter cloacae, Samples 5, 7 and 8

Negative (1) Coliform Organism - Proteus mirabilis, Sample 3

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 9

Blank - No Organism, Samples 2 and 10





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
Lab Director
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510
(707) 747-5906

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01401
B354201
03/08/13
01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080B) (Continued)

2530	Fecal Coliforms - Sample 1	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 2	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 3	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 4	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 5	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 6	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 7	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 8	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 9	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN
2530	Fecal Coliforms - Sample 10	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221E 18/19thED	2/15/2013	0			RN

Total Coliforms Evaluation : Acceptable

Fecal Coliforms Evaluation : Acceptable

E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 1, 4 and 6

Total Coliform Organism - Enterobacter cloacae, Samples 5, 7 and 8

Negative (1) Coliform Organism - Proteus mirabilis, Sample 3

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 9

Blank - No Organism, Samples 2 and 10





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
Lab Director
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510
(707) 747-5906

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01401
B354201
03/08/13
01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080B) (Continued)

2525	E.coli - Sample 1	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 2	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 3	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 4	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 5	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 6	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 7	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 8	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 9	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN
2525	E.coli - Sample 10	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9221F ECMUG online	2/15/2013	0			RN

Total Coliforms Evaluation : **Acceptable**

Fecal Coliforms Evaluation : **Acceptable**

E.coli Evaluation : **Acceptable**

Fecal Coliform Organism - Escherichia coli, Samples 1, 4 and 6

Total Coliform Organism - Enterobacter cloacae, Samples 5, 7 and 8

Negative (1) Coliform Organism - Proteus mirabilis, Sample 3

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 9

Blank - No Organism, Samples 2 and 10

All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
Lab Director
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510
(707) 747-5906

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01401
B354201
03/08/13
01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080C)

2500	Total Coliforms - Sample 1	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 2	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 3	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 4	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 5	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 6	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 7	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 8	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 9	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley
2500	Total Coliforms - Sample 10	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222B 18/19thED	2/15/2013	0			Mary Peasley

Total Coliforms Evaluation : Acceptable
Fecal Coliforms Evaluation : Not Reported
E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 5, 7 and 9
Total Coliform Organism - Enterobacter cloacae, Samples 2, 4 and 6
Negative (1) Coliform Organism - Proteus mirabilis, Sample 10
Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 1
Blank - No Organism, Samples 3 and 8





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080C) (Continued)

2530	Fecal Coliforms - Sample 1	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 2	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 3	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 4	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 5	CFU/100mL		Presence	Presence	Not Reported						
2530	Fecal Coliforms - Sample 6	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 7	CFU/100mL		Presence	Presence	Not Reported						
2530	Fecal Coliforms - Sample 8	CFU/100mL		Absence	Absence	Not Reported						
2530	Fecal Coliforms - Sample 9	CFU/100mL		Presence	Presence	Not Reported						
2530	Fecal Coliforms - Sample 10	CFU/100mL		Absence	Absence	Not Reported						

Total Coliforms Evaluation : Acceptable
Fecal Coliforms Evaluation : Not Reported
E.coli Evaluation : Acceptable

Fecal Coliform Organism - Escherichia coli, Samples 5, 7 and 9
 Total Coliform Organism - Enterobacter cloacae, Samples 2, 4 and 6
 Negative (1) Coliform Organism - Proteus mirabilis, Sample 10
 Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 1
 Blank - No Organism, Samples 3 and 8



All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
Lab Director
Biovir Laboratories
685 Stone Rd
Unit #6
Benicia, CA 94510
(707) 747-5906

EPA ID:
ERA Customer Number:
Report Issued:
Study Dates:

CA01401
B354201
03/08/13
01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080C) (Continued)

2525	E.coli - Sample 1	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 2	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 3	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 4	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 5	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 6	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 7	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 8	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 9	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley
2525	E.coli - Sample 10	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222G online	2/15/2013	0			Mary Peasley

Total Coliforms Evaluation : **Acceptable**

Fecal Coliforms Evaluation : **Not Reported**

E.coli Evaluation : **Acceptable**

Fecal Coliform Organism - Escherichia coli, Samples 5, 7 and 9

Total Coliform Organism - Enterobacter cloacae, Samples 2, 4 and 6

Negative (1) Coliform Organism - Proteus mirabilis, Sample 10

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 1

Blank - No Organism, Samples 3 and 8





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080D)

2500	Total Coliforms - Sample 1	CFU/100mL		Presence	Presence	Not Reported						
2500	Total Coliforms - Sample 2	CFU/100mL		Presence	Presence	Not Reported						
2500	Total Coliforms - Sample 3	CFU/100mL		Absence	Absence	Not Reported						
2500	Total Coliforms - Sample 4	CFU/100mL		Presence	Presence	Not Reported						
2500	Total Coliforms - Sample 5	CFU/100mL		Presence	Presence	Not Reported						
2500	Total Coliforms - Sample 6	CFU/100mL		Absence	Absence	Not Reported						
2500	Total Coliforms - Sample 7	CFU/100mL		Absence	Absence	Not Reported						
2500	Total Coliforms - Sample 8	CFU/100mL		Presence	Presence	Not Reported						
2500	Total Coliforms - Sample 9	CFU/100mL		Absence	Absence	Not Reported						
2500	Total Coliforms - Sample 10	CFU/100mL		Presence	Presence	Not Reported						

Total Coliforms Evaluation : Not Reported

Fecal Coliforms Evaluation : Acceptable

E.coli Evaluation : Not Reported

Fecal Coliform Organism - Escherichia coli, Samples 1, 4 and 5

Total Coliform Organism - Enterobacter cloacae, Samples 2, 8 and 10

Negative (1) Coliform Organism - Proteus mirabilis, Sample 9

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 6

Blank - No Organism, Samples 3 and 7





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080D) (Continued)

2530	Fecal Coliforms - Sample 1	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 2	CFU/100mL	Presence	Absence	Absence	Not Acceptable	SM9222D m FC online	2/12/2013				
2530	Fecal Coliforms - Sample 3	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 4	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 5	CFU/100mL	Presence	Presence	Presence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 6	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 7	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 8	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 9	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222D m FC online	2/12/2013	0			
2530	Fecal Coliforms - Sample 10	CFU/100mL	Absence	Absence	Absence	Acceptable	SM9222D m FC online	2/12/2013	0			

Total Coliforms Evaluation : Not Reported

Fecal Coliforms Evaluation : Acceptable

E.coli Evaluation : Not Reported

Fecal Coliform Organism - Escherichia coli, Samples 1, 4 and 5

Total Coliform Organism - Enterobacter cloacae, Samples 2, 8 and 10

Negative (1) Coliform Organism - Proteus mirabilis, Sample 9

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 6

Blank - No Organism, Samples 3 and 7



All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01





A Waters Company

WS-198 2009 TNI Evaluation Final Complete Report

Richard Danielson
 Lab Director
 Biovir Laboratories
 685 Stone Rd
 Unit #6
 Benicia, CA 94510
 (707) 747-5906

EPA ID:
 ERA Customer Number:
 Report Issued:
 Study Dates:

CA01401
 B354201
 03/08/13
 01/07/13 - 02/21/13

TNI Analyte Code	Analyte	Units	Reported Value	Assigned Value	Acceptance Limits	Performance Evaluation	Method Description	Analysis Date	Z Score	Study Mean	Study Standard Deviation	Analyst Name
------------------	---------	-------	----------------	----------------	-------------------	------------------------	--------------------	---------------	---------	------------	--------------------------	--------------

WS MicrobE™ (Coliforms) (cat# 080D) (Continued)

2525	E.coli - Sample 1	CFU/100mL		Presence	Presence	Not Reported						
2525	E.coli - Sample 2	CFU/100mL		Absence	Absence	Not Reported						
2525	E.coli - Sample 3	CFU/100mL		Absence	Absence	Not Reported						
2525	E.coli - Sample 4	CFU/100mL		Presence	Presence	Not Reported						
2525	E.coli - Sample 5	CFU/100mL		Presence	Presence	Not Reported						
2525	E.coli - Sample 6	CFU/100mL		Absence	Absence	Not Reported						
2525	E.coli - Sample 7	CFU/100mL		Absence	Absence	Not Reported						
2525	E.coli - Sample 8	CFU/100mL		Absence	Absence	Not Reported						
2525	E.coli - Sample 9	CFU/100mL		Absence	Absence	Not Reported						
2525	E.coli - Sample 10	CFU/100mL		Absence	Absence	Not Reported						

Total Coliforms Evaluation : Not Reported

Fecal Coliforms Evaluation : Acceptable

E.coli Evaluation : Not Reported

Fecal Coliform Organism - Escherichia coli, Samples 1, 4 and 5

Total Coliform Organism - Enterobacter cloacae, Samples 2, 8 and 10

Negative (1) Coliform Organism - Proteus mirabilis, Sample 9

Negative (2) Coliform Organism - Pseudomonas aeruginosa, Sample 6

Blank - No Organism, Samples 3 and 7

WS Heterotrophic Plate Count (cat# 079)

2555	Heterotrophic Plate Count (MF, PP)	CFU/mL	56	72.0	43.0 - 84.0	Acceptable	SM9215B PCA online	2/13/2013	-0.41	60.1	10.1	Mary Peasley
2555	Heterotrophic Plate Count (MPN)	MPN/mL		79.0	32.5 - 111	Not Reported				59.9	19.5	

Per Section 6.4.3 a of the 2009 TNI Standard, "The assigned values for quantitative microbiology analytes shall be equal to the mean of the assigned value verification and/or homogeneity testing per Sections 7.1 and 7.2". The final acceptance limits are derived from the calculated study mean and study standard deviation from laboratory-reported results. Disagreement between the assigned values and study means/acceptance limits are due to the inherent variability of microbiology methods and differences in the methods used by ERA and participant laboratories. For quantitative microbiology analytes, the assigned value is not used in the evaluation of laboratories.



All analytes are included in ERA's A2LA accreditation. Lab Code: 1539-01

